

Wardray Premise Limited
Northern Regional Office
Unit 2 • Springvale Works
Elland Road • Brighouse
West Yorkshire
HD6 2RN • UK
t • +44(0)1484 718000
f • +44(0)1484 718160
northern-branch@wardray-premise.com
www.wardray-premise.com

Magnetic Testing of Products at 7.0T – Standard Items

ITEM: MR4501 - Fixed Portering Chair

Manufactured by Wardray Premise Limited at: Northern Regional Office Unit 2 - Springvale Works Elland Road Brighouse West Yorkshire HD6 2RN UK

This item was tested using the 7.0T Philips Full Body Scanner at the University of Nottingham. The item was initially positioned at approximately 30cm from the mouth of the magnet bore and then gradually taken up to the mouth of the magnet from both ends and from both sides. In the case of items with braked castors, the brakes were **not** applied.

In no case was any magnetic pull detected.

Items were tested on 05.06.2006 and at the time were classified at MR Safe. However, due to the revision and clarification of the ASTM Standard F2503 in 2013 they could no longer be classified as MR Safe but reclassified as MR Conditional for use up to and including 7.0T.

A letter from the University verifying the testing is available upon request.

CE Justine Colgution

Justine Colquhoun Director

Date: 15.02.2019

ADM.19/360 Page 1 of 1